



CERTIFICATE

TEST REPORT N° 22.03679.07

According to report N° 22.03679.07, dated on 29/08/2022, we confirm that the below mentioned item was tested at CENTEXBEL with reference to ISO 10993-5 (2009) : Biological evaluation of medical devices – Part 5: Tests for in vitro cytotoxicity.

The item shows the following property

No cytotoxicity

Provided that it is properly applied.

The evaluation of the cytotoxicity is based on the standard ISO 10993-5 : 2009.

SAMPLE

Polyurethane foam: EUCASELECT® (except all Eucaselect®CM qualities (Combustion Mofified))

Company

Kabelwerk Eupen AG
Malmedyerstrasse 9
4700 EUPEN
Belgium

Certificate No. YR2884_2022 is valid until 29/08/2027

Centexbel | Technologiepark 70 | BE 9052 Gent | Belgium, 14/09/2022

Stijn Devaere, PhD - Director Services