



# CERTIFICATE

## TEST REPORT N° 22.03679.08

According to report N° 22.03679.08, dated on 29/08/2022, we confirm that the below mentioned item was tested at CENTEXBEL with reference to ISO 10993-5 (2009) : Biological evaluation of medical devices – Part 5: Tests for in vitro cytotoxicity.

The item shows the following property

## No cytotoxicity

Provided that it is properly applied.

The evaluation of the cytotoxicity is based on the standard ISO 10993-5 : 2009.

**SAMPLE**                      **Polyurethane foam: EUCAVISCO® (except EUCAVISCO® MDI and EUCAVISCO® MDI (Combustion Modified))**

**Company**                      Kabelwerk Eupen AG  
Malmedyerstrasse 9  
4700 EUPEN  
Belgium

Certificate No. YR2885\_2022 is valid until 29/08/2027

Centexbel | Technologiepark 70 | BE 9052 Gent | Belgium, 11/10/2022

Stijn Devaere, PhD - Director Services