



CERTIFICATE

CENTEXBEL TYPE TESTING | CERTIFICATION REPORT N° YR1794_2017

According to report N° 17.03336.01, 17.03336.02 and 17.03336.03 dated on 24/07/2017, we confirm that the below mentioned item was tested at CENTEXBEL with reference to Biological evaluation of medical devices – Part 5: Tests for in vitro cytotoxicity.

The item shows the following property

No cytotoxicity

Provided that it is properly applied.

The evaluation of the cytotoxicity is based on the standard ISO 10993-5 : 2009.

SAMPLE Polyurethane foam: EUCABASE®, EUCAFEEL® and EUCASELECT®

Company Kabelwerk Eupen AG
Malmedyerstrasse 9
4700 EUPEN
Belgium

Certificate No. YR1794_2017 is valid until 24/07/2022

Centexbel | Technologiepark 7 | BE 9052 Gent | Belgium, 22/11/2017


Stijn Devaere, PhD - Director Services